

Manufacturer: Beurer GmbH (see address in footer)

Product category: Infrared thermometer

Product type: FT 95

The product specified above is in conformity with the provisions of the following Union legislation and harmonized standards.

93/42/EEC Medical device directive (MDD)

UMDNS code and name: 17-888 Thermometer, Infrared, Skin

Classification/applied rule(s): Class IIa/rule 10

Conformity assessment procedure: Annex II - excluding section 4

The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate:

Certificate no. and validity: D1311700043, valid to 2024-05-26

EN 60601-1:2006 + A11:2011 + A1:2013 + A12:2014  
EN 60601-1-2:2015  
ISO 80601-2-56:2009  
EN 12470-5:2003

2014/53/EU Radio equipment directive (RED)

Draft EN 301 489-1 V2.2.0 (2017-03)  
Draft EN 301 489-17 V3.2.0 (2017-03)  
EN 55032:2015  
EN 61000-4-2:2009  
EN 61000-4-3:2006 + A1:2008 + A2:2010

2011/65/EU Restrictions of the use of certain hazardous substances in electrical and electronic equipment (ROHS)

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH

Place, date of issue: Ulm, 27.05.2020

Name, function, signature, stamp: Daniel Kämmerer, Team Leader RA

  
**Beurer GmbH**  
Söflinger Straße 218 • 89077 Ulm