

Manufacturer: Beurer GmbH (see address in footer)

Product category: Lancing device

Product type: Beurer Lancing device

The product specified above is in conformity with the provisions of the following Union legislation and harmonized standards.

93/42/EEC Medical device directive (MDD)

UMDNS code and name: 16-380 Lancing Devices, Blood

Classification/applied rule(s): Class I/rule 1

Conformity assessment procedure: Annex VII

Certificate no. and validity: D1311700042, valid to 2022-12-18

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH

Place, date of issue: Ulm, 22.06.2020

Name, function, signature, stamp: Daniel Kämmerer, Team Leader RA

Beurer GmbH
Soefflinger Straße 218 • 89077 Ulm