
	Title	EC Declaration of Conformity
	File	GL test strips GL 30_BGL 40_EC DoC_Beurer_20220208

Manufacturer:	Beurer GmbH (see address in footer)
Product category:	Blood glucose test strips for Beurer blood glucose measurement devices
Product type:	Test strips for GL 30, BGL 40
The product specified above is in conformity with the provisions of the following Union legislation and harmonized standards.	
98/79/EC	In vitro medical device directive (IVDD)
EDMA code and name:	21-06-01 Blood glucose meter
Classification:	Annex II/list B
Conformity assessment procedure:	Annex IV section 3
The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate:	
Certificate no. and validity:	D1311700050, valid to 2024-05-26
This declaration of conformity is issued under the sole responsibility of the manufacturer.	
Signed for and on behalf of:	Beurer GmbH
Place, date of issue:	Ulm, 08.02.2022
Name, function, signature, stamp:	Daniel Kämmerer, Team Leader RA  Beurer GmbH Söflinger Straße 218 • 89077 Ulm