

Manufacturer: Beurer GmbH (see address in footer)

Product category: Blood glucose test strips

Product type: Test strips GL 44 GL 44 lean GL 50 GL 50 evo

The product specified above is in conformity with the provisions of the following Union legislation and harmonized standards.

98/79/EC In vitro medical device directive (IVDD)

EDMA code and name: 21-06-01 Blood glucose meter

Classification: Annex II/list B

Conformity assessment procedure: Annex IV section 3

The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate:

Certificate no. and validity: D1311700044, valid to 2024-05-26

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH

Place, date of issue: Ulm, 24.05.2021

Name, function, signature, stamp: Werner Meternek, Director R&D / RA

 **Beurer GmbH**  
Soeflinger Straße 218 • 89077 Ulm