

Manufacturer: Beurer GmbH (see address in footer)

Product category: Lice comb

Product type: HT 15

(EU) 2017/745 Medical device regulation (MDR)

Basic-UDI-DI: 4211125BMD0020X3

CND code and name: V901001 - PEDICULOSIS ERADICATION, MECHANICAL ACTION DEVICES

Classification/applied rule(s): Class I/rule 13

Conformity assessment procedure: not applicable for class I devices

Certificate no. and validity: D1311700042, valid to 2022-12-18

EN 60601-1:2006 + A1:2013
EN 60601-1-11:2015
EN 60601-1-2:2015
IEC 60601-1-6:2010 + A1:2013
IEC 62366-1:2015 + COR1:2016
EN ISO 14971:2019
ISO 10993-1:2018
EN ISO 15223-1:2016
EN 1041:2008 + A1:2013

2011/65/EU Restrictions of the use of certain hazardous substances in electrical and electronic equipment (ROHS)

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH

Place, date of issue: Ulm, 16.09.2020

Name, function, signature, stamp: Marco Bühler, Managing Director


Beurer GmbH
Söflinger Straße 218 • 89077 Ulm