


Manufacturer: Beurer GmbH (see address in footer)  
Product category: Lice comb  
Product type: HT 15  
Intended use: The lice comb therefore provides a simple, safe and effective method to remove lice

(EU) 2017/745 Medical device regulation (MDR)  
Basic-UDI-DI: 4211125BMD0020X3  
CND code and name: V901001 - PEDICULOSIS ERADICATION, MECHANICAL ACTION DEVICES  
Classification/applied rule(s): Class I/rule 13  
Conformity assessment procedure: not applicable for class I devices  
Certificate no. and validity: D1311700048, valid to 2023-11-29  
EN 60601-1:2006 + A1:2013  
EN 60601-1-11:2015  
EN 60601-1-2:2015  
IEC 60601-1-6:2010 + A1:2013  
IEC 62366-1:2015 + COR1:2016  
EN ISO 14971:2019  
ISO 10993-1:2018  
EN ISO 15223-1:2016  
EN 1041:2008 + A1:2013

2011/65/EU Restrictions of the use of certain hazardous substances in electrical and electronic equipment (ROHS)

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH  
Place, date of issue: Ulm, 19.11.2021  
Name, function, signature, stamp: Werner Meternek, Director QM / RA  
  
**Beurer GmbH**  
Soeflinger Straße 218 • 89077 Ulm