


<b>HANS DINSLAGE GmbH</b>	<b>Title</b>	<b>EC Declaration of Conformity</b>
	<b>File</b>	SBC 26_EC DoC_Hans Dinslage_20210506

<b>Manufacturer:</b>	Hans Dinslage GmbH (see address in footer)
<b>Product category:</b>	Blood pressure monitor
<b>Product type:</b>	SBC 26
The product specified above is in conformity with the provisions of the following Union legislation and harmonized standards.	
<b>93/42/EEC</b>	Medical device directive (MDD)
<b>UMDNS code and name:</b>	16-174 Blood Pressure Monitor, electronic, manual
<b>Classification/applied rule(s):</b>	Class IIa/rule 10
<b>Conformity assessment procedure:</b>	Annex II - excluding section 4
The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate:	
<b>Certificate no. and validity:</b>	D1326900034, valid to 2024-05-26
	EN 60601-1:2006 + A1:2013 + A11:2011 + A12:2014 EN 60601-1-2:2015 IEC 80601-2-30:2009 + A1:2013
<b>2011/65/EU</b>	Restrictions of the use of certain hazardous substances in electrical and electronic equipment (ROHS)
This declaration of conformity is issued under the sole responsibility of the manufacturer.	
<b>Signed for and on behalf of:</b>	Hans Dinslage GmbH
<b>Place, date of issue:</b>	Ulm, 06.05.2021
<b>Name, function, signature, stamp:</b>	Daniel Kämmerer, Team Leader RA  <b>Hans Dinslage GmbH</b> Riedlinger Str. 28 • 88524 Uttenweiler